

Non WHO Reference Material Methylated Human Serum Albumin (mHSA) 12/176 NIBSC code: 12/176 Instructions for use (Version 3.02, Dated 18/01/2013)

This material is not for in vitro diagnostic use.

#### 1. INTENDED USE

For use as a coating reagent in Meningococcal polysaccharide ELISA assay, the mHSA solution is added in the coating buffer to a final concentration of  $5\,\mu g/ml$ .

# 2. CAUTION

# This preparation is not for administration to humans or animals in the human food chain.

The preparation contains material of human origin, and either the final product or the source materials, from which it is derived, have been tested and found negative for HBsAg, anti-HIV and HCV RNA. As with all materials of biological origin, this preparation should be regarded as potentially hazardous to health. It should be used and discarded according to your own laboratory's safety procedures. Such safety procedures should include the wearing of protective gloves and avoiding the generation of aerosols. Care should be exercised in opening ampoules or vials, to avoid cuts.

#### 3. UNITAGE

N/A

# 4. CONTENTS

Country of origin of biological material: United Kingdom.

Each ampoule contains the freeze-dried residue from 1ml of a 1mg/ml solution of methylated HSA in distilled water. The material was prepared as follows: 5g of purified human albumin was purchased from Sigma (code number A8763). The material was suspended in 500ml methanol (BDH 101586B AnalR) and 4.2ml 12N HCl (BDH 101252F) was added. The mixture was allowed to stand in the dark for 3 days with occasional mixing. The precipitate was collected by centrifugation in six tubes at 6000 rpm for 10 mins. Each of the pellets were washed 4x with 20 ml methanol (centrigfuged each time as above) and 4x with 20 ml anhydrous ether (allowed to stand for 10 mins per wash, without centrifugation). After evaporation of residual ether, the final mHSA pellet was dried under a stream of nitrogen and finally dried over potassium hydroxide pellets in a vaccuum dessicator for 10 mins. The protein concentration was estimated using the BCA method and the solution was then diluted to 1 mg/ml in distilled water. The material was filled at 1 ml per ampoule and freezedried. The final protein concentration of the material in a reconstituted ampoule was 1.4 mg/ml as measured by UV absorbance (using a theoretical absorption coefficient as described in Tarelli et al Biologicals (1998) 26, 331-346)). This concentration would be consistent when compared with previous batches of mHSA.

# 5. STORAGE

Store freeze dried ampoules and reconstituted aliquots at -20C

Please note: because of the inherent stability of lyophilized material, NIBSC may ship these materials at ambient temperature.

# 6. DIRECTIONS FOR OPENING

DIN ampoules have an 'easy-open' coloured stress point, where the narrow ampoule stem joins the wider ampoule body. Various types of ampoule breaker are available commercially. To open the ampoule, tap the ampoule gently to collect material at the bottom (labelled) end and follow manufactures instructions provided with the ampoule breaker.

# National Institute for Biological Standards and Control,

Potters Bar, Hertfordshire, EN6 3QG. T +44 (0)1707 641000, nibsc.org WHO International Laboratory for Biological Standards, UK Official Medicines Control Laboratory

#### 7. USE OF MATERIAL

No attempt should be made to weigh out any portion of the freeze-dried material prior to reconstitution

The contents of the ampoule should be rehydrated with 1 ml of sterile water. This preparation should be validated against previous batches of mHSA for individual requirements.

#### 8. STABILITY

Reference materials are held at NIBSC within assured, temperature-controlled storage facilities. Reference Materials should be stored on receipt as indicated on the label.

Assigned content of ampoule valid at time of manufacture. Comparison of the protein profiles for this and previous preparations of mHSA (99/592 and 04/142) give a similar pattern by SDS polyacrylamide gel electrophoresis. Performance of this preparation in a polysaccharide identidy ELISA was similar to the previous preparation of mHSA, 04/142. Users who have data supporting any deterioration in the characteristics of any reference preparation are encouraged to contact NIBSC.

#### 9. REFERENCES

Gheesling, L L. et al. (1994). Multicenter comparison of Neisseria meningitidis serogroup C anti-capsular polysaccharide antibody levels measured by a standardized enzyme-linked immunosorbent assay. J Clin Microbiol. 32(6):1475-82.

Carlone, GM.(1992). J Clin Microbiol. 30(1):154-9. Multicenter comparison of levels of antibody to the Neisseria meningitidis group A capsular polysaccharide measured by using an enzyme-linked immunosorbent assay.

#### 10. ACKNOWLEDGEMENTS

N/A

# 11. FURTHER INFORMATION

Further information can be obtained as follows;

This material: enquiries@nibsc.org

WHO Biological Standards:

http://www.who.int/biologicals/en/ JCTLM Higher order reference materials:

http://www.bipm.org/en/committees/jc/jctlm/

Derivation of International Units:

http://www.nibsc.org/standardisation/international\_standards.aspx

Ordering standards from NIBSC:

http://www.nibsc.org/products/ordering.aspx

NIBSC Terms & Conditions:

http://www.nibsc.org/terms\_and\_conditions.aspx

# 12. CUSTOMER FEEDBACK

Customers are encouraged to provide feedback on the suitability or use of the material provided or other aspects of our service. Please send any comments to enquiries@nibsc.org

# 13. CITATION

In all publications, including data sheets, in which this material is referenced, it is important that the preparation's title, its status, the NIBSC code number, and the name and address of NIBSC are cited and cited correctly.



#### 14. MATERIAL SAFETY SHEET

Classification in accordance with Directive 2000/54/EC, Regulation (EC) No 1272/2008: Not applicable or not classified

(EC) No 1272/2008: Not applicable or not classified		
Physical and Chemical properties		
Physical	Corrosive:	No
appearance: Freeze		
dried powder		
Stable:	Oxidising:	No
Yes		
Hygroscopic:	Irritant:	No
No		
Flammable:	Handling:	See caution, Section 2
No		
Other (specify): Contains material of human origin		
Toxicological properties		
Effects of inhalation: Not established, avoid inhalation		
Effects of ingestion: Not established, avoid ingestion		
Effects of skin absorption: Not established, avoid contact with skin		
Suggested First Aid		
Inhalation: So		
Ingestion: Seek medical advice		
Contact with eyes: W	ash with copious	amounts of water. Seek
medical advice		
Contact with skin: W	ash thoroughly wi	th water.
Action on Spillage and Method of Disposal		
Spillage of ampoule contents should be taken up with absorbent		
material wetted with an appropriate disinfectant. Rinse area with an		
appropriate disinfectant followed by water.		
Absorbent materials used to treat spillage should be treated as		

# 15. LIABILITY AND LOSS

biological waste.

In the event that this document is translated into another language, the English language version shall prevail in the event of any inconsistencies between the documents.

Unless expressly stated otherwise by NIBSC, NIBSC's Standard Terms and Conditions for the Supply of Materials (available at http://www.nibsc.org/About\_Us/Terms\_and\_Conditions.aspx or upon request by the Recipient) ("Conditions") apply to the exclusion of all other terms and are hereby incorporated into this document by reference. The Recipient's attention is drawn in particular to the provisions of clause 11 of the Conditions.

#### 16. INFORMATION FOR CUSTOMS USE ONLY

Country of origin for customs purposes\*: United Kingdom

\* Defined as the country where the goods have been produced and/or sufficiently processed to be classed as originating from the country of supply, for example a change of state such as freeze-drying.

Net weight: 4.5g.

Toxicity Statement: Toxicity not assessed

Veterinary certificate or other statement if applicable.

Attached: No

National Institute for Biological Standards and Control,

Potters Bar, Hertfordshire, EN6 3QG. T +44 (0)1707 641000, nibsc.org WHO International Laboratory for Biological Standards, UK Official Medicines Control Laboratory